



KP3 RMA Report

Probe Type: KP-3
Probe S/N: 1591
Barcode: C190883

Date Created: 11/07/2025
Tested and QC Approved by: 484
RMA: 37465

- 1.) Incoming Visual and Testing Results:
Probe came to Curwensville in the proper packaging. Customer stated probe was leaking.
- 2.) Probe Internal Inspection findings and conclusions:
Probe tested good.
- 3.) Repairs:
Cleaned and retested.
- 4.) Probe Tests results:
Meets or exceed testing specifications.

Testing Results:

Test	Results
X Center Range	0.00017 in
Y Center Range	0.00007 in
Z Repeatability	0.00018 in
Total Error while Digitizing the Ring (PTV Tolerance)	0.00112 in
Total Number of Passed Testing Runs	9

* Each test runs consist of 650 probe touches.